**SPECIFICATION** 

For

**SWITCHING POWER SUPPLY** 

M/N: MPM-S103(-C)



### 100W Medical AC / DC

| Revision His | tory          |  |  |  |
|--------------|---------------|--|--|--|
| Version      | Revise Date   | Change Items   |  |  |
| Rev. 01      | Aug. 13. 2012 | Established.   |  |  |
| Rev. 02      | Aug. 31. 2012 | Revised.   |  |  |
| Rev. 03      | Oct. 11. 2012 | Added new model: MPM-S106 which is +48V output.  |  |  |
| Rev. 04      | Dec. 6. 2012  | 1.Derating curve. 2.Dip Voltage dips 30%, 25 cycles from A to A/B.   |  |  |
| Rev. 05      | Sep. 25. 2013 | 1.Add mechanical drawing.  2.Added max. output current.  3.Efficiency up to 91%.   |  |  |
| Rev. 06      | Oct. 16. 2013 | Change derating curve.   |  |  |
| Rev. 07      | Nov. 7. 2013  | Change derating curve for MPM-S106.  |  |  |
| Rev. 08      | Feb. 18. 2014 | Added optional cover kit drawing and its derating curve.   |  |  |
| Rev. 09      | Jul. 1. 2015  | Added TUV logo.  |  |  |
| Rev. 10      | Jul. 22. 2015 | Changed Molex Housing part no.   |  |  |
| Rev. 11      | Aug. 27. 2015 | Changed Safety Approvals for UL approved.  |  |  |
| Rev. 12      | Sep. 10. 2015 | Changed IEC 60601-1: 2005 3rd Edition for UL approved.   |  |  |
| Rev. 13      | Nov. 25. 2015 | 1.Added "or equivalent" after "Molex" and "JST".  2.Changed Molex Proposed Terminals from 5176 to 5167.  3.Added vibration test. |  |  |
| Rev. 14      | Jan. 23. 2017 | 1.Added "Designed to meet IEC 60601-1-2 4th ed. EMC".  2.Changed IEC 61000-4-11 Voltage interruptions >95%, 250 cycles to C.     |  |  |
| Rev. 15      | Apr. 19. 2017 | Added MOPP x 2 isolation.  |  |  |
| Rev. 16      | May 25. 2017  | Added MPM-S103-C performance curve at 70 degree C.   |  |  |
| Rev. 17      | Dec. 21.2017  | 1.Added performance curve with case.     2.Changed Form.   |  |  |
| Rev. 18      | Mar. 9. 2018  | 1.Added Designed to meet IEC 60601-1-2 4th ed. EMC.     2.Changed EMC and Safety Approvals.                                      |  |  |
| Rev. 19      | Jul. 3. 2018  | Changed mechanical diagram.  |  |  |
| Rev. 20      | Nov. 6. 2018  | 1.Changed EMC: Immunity ESD to ±15KV air discharge, ±8KV contact discharge.     2.Changed EMC: Immunity Power Magnetic to 30A/m. |  |  |
| Rev. 21      | Jun. 11. 2019 | Changed safety standard for IEC , EN and UL/cUL.   |  |  |
| Rev. 22      | May. 27. 2022 | Changed "IP to Ground" to 1800VAC.   |  |  |























### **FEATURES**

- 100W with convection-cooled single output power supply.
- High efficiency up to 91%.
- No-load power consumption < 0.5W.
- Class II design, additional class I functional ground
- Compact size 2 x 4 inch and low profile.
- Medical standard EN / UL 60601-1 3.1 Edition approved, MOPP x 2 isolation.
- IEC 60601-1-2 4th ed. EMC approved.

**Input Connector Type** 

- Meets EMI CISPR/FCC class B.
- Optional cover kit.

### **Models & Ratings**

| Model Number | Wattage<br>(Rated / Max) | Output Voltage | Min. Current | Rated<br>Current | Max. Current |
|--------------|--------------------------|----------------|--------------|------------------|--------------|
| MPM-S103     | 100 W / 115 W            | +12 V          | 0 A          | 8.4 A            | 9.6 A        |

Total Output Power: Max. 100W convection cooled, above 101~115W with 7 CFM forced air-cooling at 50°C environment temperature. Please see detail performance curves as below. Note:

1.Model no. coding:

### MPM-S10











(2)

|       | Molex Type Connector<br>or equivalent |
|-------|---------------------------------------|
| blank |                                       |









Please refer to paragraph 8 for detail.

| 1 |   | ` |
|---|---|---|
|   | 2 |   |
| ( | J | , |

| Z =   | Optional cover |
|-------|----------------|
| Blank | Open fame      |
| С     | With optional  |
|       | cover kit      |

+12

#### **Summary**

| Characteristic        | Minimum        | Typical   | Maximum | Units | Notes & Conditions  |  |  |
|-----------------------|----------------|---|---------|-------|---|--|--|
| Input Range           | 85             | 115 / 230   | 264     | VAC   | Universal input range.  |  |  |
| Input Frequency       | 47             | 50 / 60   | 63      | Hz    | AC input.   |  |  |
| Efficiency            |                | 90  |         | %     | At input 230VAC, rated load, 0.5 hr. warm up.                                 |  |  |
| Operation Temperature | -20            |   | +70     | °C    | Derate linearly above 50°C, please refer to the following performance curves. |  |  |
| Weight                |                | 154.2   |         | g     |   |  |  |
| Dimensions            | 101.6 (L) x 50 | 101.6 (L) x 50.8 (W) x 33.0 (H) mm, Tolerance +/- 0.4mm.  |         |       |   |  |  |
| EMC                   |                | EN 55011 / CISPR 11 & FCC Part 18, EN 60601-1-2, EN 61000-3-2, EN 61000-3-3, IEC 61000-4-3, IEC 61000-4-4, IEC 61000-4-5, IEC 61000-4-6, IEC 61000-4-8, IEC 61000-4-11  |         |       |   |  |  |
| Safety Approvals      | UL/cUL: ANSI   | IEC 60601-1: 2005+A1:2012,3.1 Edition, EN 60601-1: 2006+A11+A1+A12, 3.1 Edition, UL/cUL: ANSI/AAMI ES60601-1(2005/(R)2012 + A1:2012, C1:2009/(R)2012 + A2:2010/(R)2012), CAN/CSA-C22.2 No. 60601-1:14 - 3.1 Edition |         |       |   |  |  |

# MPM-S103(-C)

### 100W Medical AC / DC

| Input                     |               |  |         |       |  |  |  |
|---------------------------|---------------|--|---------|-------|--|--|--|
| Characteristic            | Minimum       | Typical  | Maximum | Units | Notes & Conditions   |  |  |
| Input Voltage             | 85            | 115 / 230  | 264     | VAC   | Universal input range.   |  |  |
| Label Voltage             | 100           |  | 240     | VAC   |  |  |  |
| Input Frequency           | 47            | 50 / 60  | 63      | Hz    | AC input.  |  |  |
| Input Current             |               |  | 2 / 1.2 | А     | Nominal AC Input Voltage (115VAC/230VAC), rated load.                      |  |  |
| Inrush Current            |               |  | 30 / 60 | А     | Nominal AC Input Voltage (115VAC/230VAC), one cycle at 25°C cold start.    |  |  |
| Leakage Current           |               | 100 / 300  |         | ^     | Primary to Secondary<br>Normal Condition / Single Fault Condition          |  |  |
| Leakage Current           |               | 100 / 300  |         | μΑ    | Primary to Earth GND (Note 1)<br>Normal Condition / Single Fault Condition |  |  |
| No-load power consumption |               |  | <0.5    | W     | Nominal AC Input Voltage (115VAC/230VAC).                                  |  |  |
| Input Protection          | Dual non-user | Dual non-user serviceable internally located AC input line fuse. Fuse: 3.15A / 250VAC * 2pcs |         |       |  |  |  |

Note:

<sup>1.</sup> Only exist when earth ground was connecting.

| Output                   |                 |  |                 |                   |   |  |  |
|--------------------------|-----------------|--|-----------------|-------------------|---|--|--|
| Characteristic           | Minimum         | Typical  | Maximum         | Units             | Notes & Conditions  |  |  |
| Output Voltage           |                 | 12   |                 | VDC               |   |  |  |
| Output Current           |                 | 8.4  | 9.6             | Α                 |   |  |  |
| Initial Set Accuracy     | 11.8            |  | 12.2            | V                 | Initial setting accuracy is adjusted at input 115VAC and output at 60% rated load.  |  |  |
| Minimum Load             |                 | 0  |                 | Α                 |   |  |  |
| Start Up Delay           |                 | 0.7  |                 | Sec               | Nominal AC Input Voltage (115VAC/230VAC), rated load at 25°C. Time required for initial output voltage stabilization.   |  |  |
| Hold Up Time             | 12 / 20         |  |                 | mS                | Nominal AC Input Voltage (115VAC/230VAC), rated load.   |  |  |
| Line Regulation          |                 | ±1.0   |                 | %                 | Less than ±1% at rated load with ±10% changing in input voltage.  |  |  |
| Load Regulation          |                 | ±1.0   |                 | %                 | Measured from 60% to 100% rated load and from 60% to 20% rated load (60% ±40% rated load).  |  |  |
| Ripple & Noise           |                 | 120  |                 | mV                | Measured at rated road by a 20MHz bandwidth limited oscilloscope and each output is connected with a 10μF Electrolytic Capacitor and a 0.1μF Ceramic Capacitor. |  |  |
| Overvoltage Protection   |                 | For some reason the power supply fails to control itself, the build-in over voltage protection circuit will auto recovery the outputs to prevent damaging external circuits. |                 |                   |   |  |  |
| Short Circuit Protection | Fully protected | against output   | overload and sh | ort circuit. Auto | omatic recovery upon of overload condition.   |  |  |




| Genera     | al           |         |         |         |       |   |
|------------|--------------|---------|---------|---------|-------|---|
| Cha        | aracteristic | Minimum | Typical | Maximum | Units | Notes & Conditions                            |
| Efficiency |              |         | 90      |         | %     | At input 230VAC, rated load, 0.5 hr. warm up. |
| laalation  | IP to OP     | 4000    |         |         | VAC   | 2 x MOPP                                      |
| Isolation  | IP to Ground | 1800    |         |         | VAC   | 2 x MOPP (Screw head to primary)              |
| Switching  | Frequency    |         | <65     |         | KHZ   |   |

| Environmental                         |         |             |         |       |  |
|---------------------------------------|---------|-------------|---------|-------|--|
| Characteristic                        | Minimum | Typical     | Maximum | Units | Notes & Conditions   |
| Low temperature start up              | -40     |             |         | °C    | Some specification parameters maybe exceeded until after 20 minutes warm up period. (Note 1)   |
| Operating Temperature                 | -20     |             | +70     | °C    | Derate linearly above 50°C, please refer to the following performance curves.  |
| Storage Temperature                   | -40     |             | +85     | °C    |  |
| Relative Humidity                     | 5       |             | 95      | %RH   | Non-condensing.  |
| Cooling                               | 7       |             |         | CFM   | Forced-cooled > 115W   |
| Operating /<br>Non-Operating Altitude |         | 3000 / 4000 |         | m     |  |
| Vibration                             | 0.26    |             | 6.09    | G     | Frequency Type: Sweep Frequency Frequency Range: 10~55 Hz Displacement: 1.0mm Sweep Rate: 60 minute / cycle Number of cycle: 1 cycle / axis Direction: X ,Y and Z axis |

#### Note:


### **Derating curve**







<sup>1.</sup> To start up unit, the output power should be derated to 20% rated load @ Vin < 115VAC, or derated to 40% rated load @ Vin < 230VAC, and don't need derated @ Vin ≥ 230VAC.

### 100W Medical AC / DC







# MPM-S103(-C)

### 100W Medical AC / DC

### **EMC: Emissions**

| Phenomenon       | Standard  | Class | Notes & Conditions |
|------------------|---|-------|--------------------|
| Conducted        | EN 55011 / CISPR 11 & FCC Part 18, EN 60601-1-2 | В     |                    |
| Radiated         | EN 55011 / CISPR 11 & FCC Part 18, EN 60601-1-2 | В     |                    |
| Harmonic Current | EN 61000-3-2                                    | Α     |                    |
| Voltage Flicker  | EN 61000-3-3                                    |       |                    |

#### Note:

- 1. As a build-in type power supply, the power supply needs to be installed in a suitable enclosure to pass the EMI/EMC tests. The final assembly has to comply with the valid EMI/EMC and safety.
- 2. The mounting holes should be connected to each other to conforming the EMI limit.
- 3. Apply to output equal or below 100W, for higher output power, please re-confirm with us.

### **EMC: Immunity**

| Phenomenon             | Standard       | Criteria                 | Notes & Conditions   |
|------------------------|----------------|--------------------------|--|
| ESD                    | IEC 61000-4-2  | А                        | ±15KV air discharge, ±8KV contact discharge  |
| Radiated               | IEC 61000-4-3  | А                        | 10V/m, 80 - 2700MHz  |
| EFT                    | IEC 61000-4-4  | Α                        | ±2KV Line & PE, 100KHz   |
| Surges                 | IEC 61000-4-5  | Α                        | L-N:±1KV, L/N-PE:±2KV  |
| Conducted              | IEC 61000-4-6  | Α                        | 10Vrms   |
| Power Magnetic         | IEC 61000-4-8  | Α                        | 30A/m  |
| Dips and Interruptions | IEC 61000-4-11 | A<br>A / B<br>A / B<br>C | DIP: >95%, 0.5 cycle<br>DIP: 30%, 25 cycles (Note 2)<br>DIP: 60%, 5 cycles (Note 2)<br>INT: >95%, 250 cycles |

#### Note:

- 1. Apply to output equal or below 100W, for higher output power, please re-confirm with us.
- 2. The test result of input 240Vac / 100Vac is criteria A / B.


### **Safety Approvals**

| Safety Agency | Safety Standard   | Notes & Conditions |
|---------------|---|--------------------|
| TUV           | EN 60601-1: 2006+A11+A1+A12, 3.1 Edition  | Designed to meet.  |
| СВ            | IEC 60601-1: 2005+CORR.1:2006+CORR.2: 2007+A1:2012, 3.1 Edition   | Approved.          |
| UL/cUL        | ANSI/AAMI ES60601-1(2005/(R)2012 + A1:2012, C1:2009/(R)2012 + A2:2010/(R)2012), CAN/CSA-C22.2 No. 60601-1:14 -3.1 Edition | Approved.          |



### **Mechanical Details**


 $\label{eq:SIZE} SIZE: 101.6(L)~x~50.8(W)~x~33.0(H)mm,~Tolerance~+/-0.4mm.$ 





MPM-S103-C

108.7 (L) x 58.0 (W) x 38.4 (H) mm, Tolerance +/- 0.4mm.



| Parameter                  | Conditions/Description |  |            |                                    |                                 |  |  |
|----------------------------|------------------------|--|------------|------------------------------------|---------------------------------|--|--|
| Dimension                  | 101.6 (L) x            | 101.6 (L) x 50.8 (W) x 33.0 (H) mm, Tolerance +/- 0.4mm. |            |                                    |                                 |  |  |
| Connector & Pin Assignment | Location               | Pin  | Assignment | Proposed Housing                   | Proposed Terminals              |  |  |
|                            |                        | 1  | AC in (L)  | MOLEX: 09-50-1031 (5195-03) or     | MOLEX: 5194 or 5225             |  |  |
|                            | CN1                    |  |            | 09-52-4034 (5239-03) or            | 2478, 2578,5167 or 5168 or      |  |  |
|                            | (Input)                | 2 AC in (N)  | AC in (N)  | equivalent;                        | equivalent;                     |  |  |
|                            |                        |  |            | JST: VHR-3N (Note 1) or equivalent | JST: SVH-21T-P1.1 or equivalent |  |  |
|                            |                        | 1  | 0 V        | MOLEX: 09-50-1041 (5195-04) or     | MOLEX: 5194 or 5225             |  |  |
|                            | CN2                    | 2  | 0 V        | 09-52-4044 (5239-04)or             | 2478, 2578,5167 or 5168 or      |  |  |
|                            | (Output)               | 3  | + V        | equivalent;                        | equivalent;                     |  |  |
|                            |                        | 4  | + V        | JST: VHR-4N (Note 1) or equivalent | JST: SVH-21T-P1.1 or equivalent |  |  |

Note: 1. Exist with model no. suffixed -J, please see the Model no. coding.



## MPM-S103(-C)

### 100W Medical AC / DC

#### **Thermal Considerations**

In order to ensure safe operation of the PSU in the end-use equipment, the temperature of the components listed in the table below must not be exceeded.

Temperature should be monitored using J type thermocouples placed on the hottest part of the component (out of any direct air flow). See Mechanical Details for component locations.

| Temperature Measurements at max. amb. |                 |  |  |  |  |
|---------------------------------------|-----------------|--|--|--|--|
| Component                             | Max Temperature |  |  |  |  |
| T1                                    | 110°C           |  |  |  |  |
| Q1                                    | 120°C           |  |  |  |  |
| D6                                    | 120°C           |  |  |  |  |
| C5                                    | 105℃            |  |  |  |  |
| C9                                    | 105°C           |  |  |  |  |

