



**QQHM2.E194574
Power Supplies, Medical and Dental - Component**

Page Bottom

Power Supplies, Medical and Dental - Component

See General Information for Power Supplies, Medical and Dental - Component

MAGIC POWER TECHNOLOGY CO LTD

E194574

9TH FL-6

268 LIEN CHENG RD

CHUNG-HO, TAIPEI 235 TAIWAN

Model No.	Rated Input			Max Output				OC	SP	EP	FC	GC
	Volts	Hz	SC	V	A	VA						
ADN-9208[*r]												
	100-240ac	50/60	0	3.3	4.5	-	9	2601-1	20B	0	0	
				5.1	6.5	-	9					
				8	3.5	-	9					
				12	2.5	-	9					
				15	1	-	9					
				24	3	-	9					
				-12	1	-	9					
				-15	1	-	9					
DTS-2420[*r]												
	100-120, 230ac	50-60	0	25	1.7	-	9	2601-1	20B	0	1	
MEM-025X[*r]												
	100-240ac	50/60	0	3.3	6	-	9	60601-1	20B	0	0	
				5.1	9	-	9					
				8	3.5	-	9					
				12	5	-	9					
				-12	1	-	9					
				15	1	-	9					
				-15	1	-	9					
MPM-60-05, MPM-60-05C[*r]												
	100-240ac	50-60	0	5	12	60	3	60601-1	20B	0	0	
MPM-60-12, MPM-60-12C[*r]												
	100-240ac	50-60	0	12	5	60	3	60601-1	20B	0	0	
MPM-60-15, MPM-60-15C[*r]												
	100-240ac	50-60	0	15	4	60	3	60601-1	20B	0	0	
MPM-60-24, MPM-60-24C[*r]												
	100-240ac	50-60	0	24	2.7	64.8	3	60601-1	20B	0	0	
MPM-815H[*r]												
	100-240ac	47-63	0	3.3	12	-	9	60601-1	20B	0	1	
				5	14	-	9					
				12	10	-	9					

				-12	1	-	9				
				5 (sb)	1.5	-	9				
MPM-842P, MPM-842P-F[*r]											
	100-240ac	47-63	0	3.3	22	-	9	60601-1	20B	0	1
				5	21	-	9				
				12	22	-	9				
				-12	0.8	-	9				
				5 (sb)	1.5	-	9				
MPM-9561[*r]											
	100-250ac	47-63	0	12	5.0	30	4	2601-1	20B	0	0
				-12	5.0	30	4				
MPM-9563[*r]											
	100-250ac	47-63	0	12	5.0	36	4	2601-1	20B	0	0
MPM-9566[*r]											
	100-250ac	47-63	0	5	12	60	4	2601-1	20B	0	0
MPM-9567											
	100-250ac	47-63	0	12	5.5	66	4	2601-1	20B	0	0
MPM-9568											
	100-250ac	47-63	0	15	4.5	68	4	2601-1	20B	0	0
MPM-9569											
	100-250ac	47-63	0	24	2.75	66	4	2601-1	20B	0	0
MPM-A30H, MPM-A30H-T											
	100-240ac	47-63	0	3.75	32.06	95.63	9	60601-1	20B	0	1
				11.99	26.01	307.84	9				
				5.10	30.5	150.98	9				
				5.01	2.91	13.66	9				
				12.16	2.15	24.21	9				
				5.04 (sb)	2.0	9.48	9				
MPM-X123[*r]											
	100-240ac	47-63	0	12-13.8	10-8.69	120	9	60601-1	20B	9	1
MPM-X124-1[*r]											
	100-240ac	47-63	0	19-20	6.3-6.0	120	9	60601-1	20B	9	1
MPM-X125[*r]											
	100-240ac	47-63	0	36	3.33	120	9	60601-1	20B	9	1
MPM-X125-1[*r]											
	100-240ac	47-63	0	36	3.33	120	9	60601-1	20B	9	1
MPM-X50-05[*r]											
	100-240ac	50-60	0	5dc	7	35	9	60601-1	20B	2	1
MPM-X50-12[*r]											
	100-240ac	50-60	0	12dc	4.17	50.04	9	60601-1	20B	2	1
MPM-X50-138[*r]											
	100-240ac	50-60	0	13.8dc	3.62	49.96	9	60601-1	20B	2	1
MPM-X50-15[*r]											
	100-240ac	50-60	0	15dc	3.33	49.95	9	60601-1	20B	2	1
MPM-X50-16[*r]											
	100-240ac	50-60	0	16dc	3.13	50.08	9	60601-1	20B	2	1

MPM-X50-18[*r]											
	100-240ac	50-60	0	18dc	2.77	49.97	9	60601-1	20B	2	1
MPM-X50-19[*r]											
	100-240ac	50-60	0	19dc	2.63	49.97	9	60601-1	20B	2	1
MPM-X50-24[*r]											
	100-240ac	50-60	0	24dc	2.1	50.4	9	60601-1	20B	2	1
MPM-X50-48[*r]											
	100-240ac	50-60	0	48dc	1.04	49.92	9	60601-1	20B	2	1

[*r] - Output values are rated.

Marking: Company name and model designation.
 Last Updated on 2009-02-03

[Questions?](#)

[Notice of Disclaimer](#)

[Page Top](#)

Copyright ©2009 Underwriters Laboratories Inc.®

The appearance of a company's name or product in this database does not in itself assure that products so identified have been manufactured under UL's Follow-Up Service. Only those products bearing the UL Mark should be considered to be Listed and covered under UL's Follow-Up Service. Always look for the Mark on the product.

UL permits the reproduction of the material contained in the Online Certification Directory subject to the following conditions: 1. The Guide Information, Designs and/or Listings (files) must be presented in their entirety and in a non-misleading manner, without any manipulation of the data (or drawings). 2. The statement "Reprinted from the Online Certifications Directory with permission from Underwriters Laboratories Inc." must appear adjacent to the extracted material. In addition, the reprinted material must include a copyright notice in the following format: "Copyright © 2009 Underwriters Laboratories Inc.®"

An independent organization working for a safer world with integrity, precision and knowledge.

